



ILLUMINATED MICROSCOPE INSTRUCTIONS



1. Eyepiece
2. Battery cover
3. Focusing Wheel
4. Light Condenser
5. Light Switch
6. Bulb
7. Transparent Cover


OPERATION

1. Place the transparent cover (7) on the object.
2. Position the hole over the area to be viewed.
3. Slide the light switch (5) to the ON position and slide down the light condenser (4) to concentrate the light on the object being viewed.
4. Look through the eye piece (1) and turn the focus wheel (3) until the object is in sharp focus.
5. When viewing is completed, turn the light switch (5) to the OFF position and retract the light condenser (4).




BATTERY REPLACEMENT

1. Remove the battery cover (2) by sliding it in the direction indicated by the arrow.
2. Remove the exhausted batteries and replace with two new AA batteries.
3. Replace the battery cover.



BULB REPLACEMENT

1. Remove the transparent cover (7) by sliding it in the direction shown in the illustration. (Note: The light condenser (4) must be retracted)
2. Unscrew the exhausted bulb and replace with a new one.
3. Replace the transparent cover.



WARRANTY, SERVICE, OR RECALIBRATION

For warranty, service, or recalibration, contact:

CONTROL COMPANY

4455 Rex Road
Friendswood, Texas 77546 USA
Ph. 281-482-1714 Fax 281-482-9448
E-mail sales@control3.com
www.control3.com

Control Company is ISO 9001 Quality-Certified by DNV and ISO 17025 accredited as a Calibration Laboratory by A2LA.