TRACEABLE® TWO-MEMORY TIMER INSTRUCTIONS

OPERATING INSTRUCTION:

To Set Clock

- 1) 12 Hour AM/PM format or 24 Hour International time user-selectable. Slide switch to CLOCK. Press 12HR \rightleftharpoons 24HR button for desired format.
- 2) Depress HR, MIN, SEC buttons to actual time. Slide switch to LOCK mode after setting.

Count-up (The lower part of display)

In CLOCK mode, press START once to start count-up function, second press is to pause. Third press to restart. Press CLEAR after pausing will reset digits to 00:00:00.

To Set Timer

- 1. Slide switch to T1 for TIMER 1, T2 for TIMER 2. After selecting which timer is to set, press HR, MIN, SEC for desired timing period. Press MEMORY afterwards (The word "MEMORY" will display) will enter the timing period into memory. Press CLEAR after MEMORY is pressed, the digits will be reset to 00:00:00. Recall memory before setting timing period (press MEMORY in T1 or T2 mode), saving resetting effort of the same timing cycle. After setting timing period, slide switch to LOCK mode.
- Press START store to start timer, second press to pause timer, third press is to restart.
- Press CLEAR after pausing will reset digits to 00:00:00. The timer will count up from 00:00:00 to tell you how long since the alarm sounded (the timing cycle is ended) To stop alarm sound or count-up, press START STOP
- 4. You can have both timer working simultaneously. After setting different timing cycle for each timer and start timing, slide switch to T 1 2 to have both timers display.
- * Count-up or timer count-down functions can only be started in LOCK mode.

Warranty, Service, or Recalibration

For warranty, service, or recalibration, contact:

CONTROL COMPANY

4455 Rex Road
Friendswood, Texas 77546 USA
Ph. 281-482-1714 Fax 281-482-9448
E-mail sales@control3.com
www.control3.com

Control Company is ISO 9001 Quality-Certified by DNV and ISO 17025 accredited as a Calibration Laboratory by A2LA.

Rev. 1 010308