

SPECIFICATIONS

Display: LCD, 1/2" digits

Range: -58 to 572°F (-50 to 300°C)

Resolution: 0.1° from -20 to 200°, 1° outside this range

Accuracy: ±1°C between -20 and 100°C

Power: G-13 silver-oxide battery

Size: 4 5/8 x 2 7/8 x 1 inches


Accessories

Supplied: Stainless steel probe, battery, stand, clip, Traceable® Certificate, and instructions.

DESCRIPTION

1. Display
2. °C/°F Button
3. On/Off Button
4. Memory Button
5. H/T Button
6. Light Button
7. Max/Min Button
8. Alarm Off Button
9. Recall Button
10. Clear Button
11. Low Alarm Button
12. High Alarm Button

Fig. 1



OPERATION

1. Turn on the unit by pressing the On/Off Button (3, Fig. 1) once.
2. Press the °F/°C Button to select the desired temperature unit. (2, Fig. 1)
3. Immerse the probe tip at least 10 millimeters into the desired substance to obtain temperature reading.
4. Press the H/T button (5, Fig. 1) once to "freeze" the display at the current temperature reading. "HOLD" flashes in the upper portion of the display indicating that the unit is in hold mode. Press the H/T button again to release the reading.
5. Press the MAX/MIN button (7, Fig. 1) once to recall the minimum temperature. "MIN" will appear on the display.
6. Press the MAX/MIN button once again to recall the maximum temperature. "MAX" will appear on the display.
7. After 3 seconds, the unit will automatically return to the current temperature and "Min" or "Max" will no longer appear on the display.
8. To back light the display, press the LIGHT button (6, Fig. 1). The display will light for 3 seconds.
9. The unit will automatically turn itself off after one hour to conserve power. To turn off this function, press Clear (10, Fig. 1) and On/Off (3, Fig. 1). To restore this function, turn the unit off and back on again by pressing the On/Off button two times.

10. A stand is located on the back of the unit. Simply pull the bottom of the stand away from the unit to extend the stand for use. Press the stand back into the unit until it snaps to close. Also located on the back of the unit is a clip.
11. With the stand extended, pull it gently away from the unit to expose the battery compartment and probe cable.

SETTING THE ALARM

1. Press and hold the Low Alarm Button (11, Fig. 1) until the desired temperature is reached. "LOW" will appear in the display. When the current temperature is at or below this set value, an alarm will beep.
2. The alarm will continue to beep four times every minute until the Alarm Off Button (8, Fig. 1) is pressed or the current temperature returns to an "in range" reading.
3. Press and hold the High Alarm Button (12, Fig. 1) until the desired temperature is reached. "HIGH" will appear in the display. When the current temperature is at or above this set value, an alarm will beep.
4. When both a High and Low alarm temperature are set, "HI LOW" will appear in the lower portion of the display.
5. To clear the alarms, press the Alarm Off button (8, Fig. 1).

MEMORY

1. Press the H/T button (5, Fig. 1) to hold the current temperature. Press the Memory button (4, Fig. 1) to store the temperature reading into memory. "Mem 1" will appear on the display. Repeat these steps to store additional temperatures. Up to 10 readings can be stored.
2. To recall a stored reading, press the Recall Button (9, Fig. 1) once. The temperature stored in Memory One will be displayed. Each subsequent press of the Recall Button will show the next stored reading.
3. To clear all stored memories, press and hold the Clear Button (10, Fig. 1) for 5 seconds.

LOW BATTERY

Erratic readings, a faint display, or no display are all indicators that the battery is low and needs to be replaced. To remove the battery, using a small screwdriver, unscrew the three screws located on the back of the thermometer inside the probe compartment. Remove the battery cover. Remove the exhausted battery and replace it with a new 357 size battery. Make certain the positive (+) side is visible. Replace the battery cover. Screw the three screws back into the unit. Do not over-tighten the screws as this may cause damage to the battery cover. Close the probe compartment.

CABLE COMPARTMENT

To expose more of the 32" cable, simply slide down the battery compartment on the back of the unit. Unwind the stored cable and replace battery compartment cover, allowing cable to rest in the opening provided.

ALL OPERATIONAL DIFFICULTIES

If this thermometer does not function properly for any reason, please properly replace the battery with a new battery (see Low Battery section, above). Low battery power can occasionally cause any number of "apparent" operational difficulties. Replacing the battery with a new fresh battery will solve most difficulties.

**WARRANTY, SERVICE, OR
RECALIBRATION**

For warranty, service, or recalibration, contact:

TRACEABLE® PRODUCTS

12554 Old Galveston Rd. Suite B230
Webster, Texas 77598 USA
Ph. 281 482-1714 • Fax 281 482-9448
E-mail support@traceable.com
www.traceable.com

Traceable® Products is ISO 9001:2015 Quality-
Certified by DNV and ISO/IEC 17025:2005
accredited as a Calibration Laboratory by A2LA.

**TRACEABLE®
MEMORY/
WATERPROOF
THERMOMETER
INSTRUCTIONS**

Traceable® is a registered trademark of Cole-Parmer.

©2018 Traceable® Products. 92-4373-00 Rev. 6 060618