

# **INITIAL SETUP**

Remove battery cover and insert batteries.



Once batteries are installed, press O button to select setting.



The first setting, Hours, will flash on display.




Press "∧" and "∨" buttons to adjust setting.




Once finished, press O button to save setting and proceed to Minutes, 12/24 Hour, etc.




- 6 Complete all settings to configure device:
  - Hours
  - Minutes
  - 12 or 24 hour format clock
  - Temperature units





### **DISPLAY DETAILS**



### CLEARING INDIVIDUAL READINGS

- 1. Press select button until desired reading to be cleared is flashing.
- 2. Press and hold clear/revert button for 3 seconds to clear reading.

### **CLEARING ALL MEMORY READINGS**

1. Press and hold clear/revert button for 3 seconds to clear all memory readings.

## WARRANTY, SERVICE, OR RECALIBRATION

# Traceable<sup>®</sup>

12554 Old Galveston Rd. Suite B230 Webster, Texas 77598 USA Ph. 281-482-1714 • Fax 281-482-9448 support@traceable.com traceable.com

Traceable® Products are ISO 9001:2018 Quality-Certified by DNV and ISO/IEC 17025:2017 accredited as a Calibration Laboratory by A2LA.

Item No. 56000-10 Model No. 5660 ©2022 1065T6\_M\_92-5660-00 Rev. 0 09022022

# Traceable® Thermohygrometer with Clock Quick-Start Guide

